



Entopsis  
 7600 NW 69<sup>th</sup> Ave  
 Medley, FL 33166  
 info@entopsis.com  
 Ph#: 1-888-407-5070

## Certificate of Analysis - Reagent RVD-E

<b><u>Product Name:</u></b>	Reagent RVD-E	<b><u>Date of Manufacture:</u></b>	November 2020
<b><u>Catalog #:</u></b>	7833025 / 7833100 / 7833500	<b><u>Lot #:</u></b>	0003
<b><u>Product Line:</u></b>	PCRopsis™	<b><u>Expiration Date:</u></b>	May 2022
		<b><u>Storage Conditions:</u></b>	4 - 25°C
		<b><u>Handling Conditions:</u></b>	Keep sterile

Testing Characteristic	Test Result	Lot
Compare Ct values between Reagent RVD-E + RT-qPCR to RNA extraction + RT-qPCR from stock sample	Observed Ct relative to RNA extraction: <b>+3.3</b>	Pass: +/- 5Ct: <b>PASS</b>
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C	No Growth	<b>PASS</b>

**Date of Quality Control Testing:** November 2020

This certificate of analysis verifies the above-mentioned lot of Reagent RVD-E conforms to intended quality parameters on the indicated date above.

**For In Vitro Diagnostic Use.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto  
 \_\_\_\_\_  
 Name of Approving Manager

  
 \_\_\_\_\_  
 Signature of Approving Manager