



Entopsis, Inc.
7600 NW 69th Ave
Medley, FL 33166
info@entopsis.com
Ph#: 1-888-407-5070

Certificate of Analysis - Reagent RVD with RVD Enhancer

<u>Product Name:</u>	Reagent RVD with RVD Enhancer	<u>Date of Manufacture:</u>	January 2022
<u>Catalog #:</u>	78336001 / 78336025 / 78336100	<u>Lot #:</u>	0022
<u>Product Line:</u>	PCRopsis TM	<u>Expiration Date:</u>	July 2023
		<u>Storage Conditions:</u>	4 - 25°C
		<u>Handling Conditions:</u>	Keep sterile

Testing Characteristic	Test Result	Lot
Compare Ct values between Reagent RVD + RT-qPCR to RNA extraction + RT-qPCR from stock sample	Observed Ct relative to RNA extraction: -1.22	Pass: +/- 5Ct: PASS
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C	No Growth	PASS

Date of Quality Control Testing: January 2022

This certificate of analysis verifies the above-mentioned lot of Reagent RVD with RVD Enhancer conforms to intended quality parameters on the indicated date above.

For In Vitro Diagnostic Use. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager



Signature of Approving Manager