



Entopsis, Inc.  
7600 NW 69<sup>th</sup> Ave  
Medley, FL 33166  
info@entopsis.com  
Ph#: 1-888-407-5070

## Certificate of Analysis - Reagent RVD with RVD Enhancer

<b><u>Product Name:</u></b>	Reagent RVD with RVD Enhancer	<b><u>Date of Manufacture:</u></b>	January 2022
<b><u>Catalog #:</u></b>	78336001 / 78336025 / 78336100	<b><u>Lot #:</u></b>	0022
<b><u>Product Line:</u></b>	PCRopsis <sup>TM</sup>	<b><u>Expiration Date:</u></b>	July 2023
		<b><u>Storage Conditions:</u></b>	4 - 25°C
		<b><u>Handling Conditions:</u></b>	Keep sterile

### Testing Characteristic

### Test Result

### Lot

Compare Ct values between Reagent RVD + RT-qPCR to RNA extraction + RT-qPCR from stock sample

Observed Ct relative to RNA extraction:  
**-1.22**

Pass: +/- 5Ct:  
**PASS**

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

**PASS**

**Date of Quality Control Testing:** January 2022

This certificate of analysis verifies the above-mentioned lot of Reagent RVD with RVD Enhancer conforms to intended quality parameters on the indicated date above.

**For Research Use Only.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager