



Entopsis, Inc.  
7600 NW 69<sup>th</sup> Ave  
Medley, FL 33166  
info@entopsis.com  
Ph#: 1-888-407-5070

## Certificate of Analysis - Reagent RVD with RVD Enhancer

**Product Name:** Reagent RVD with RVD Enhancer  
**Catalog #:** 78336001 / 78336025 / 783361000  
**Product Line:** PCROpsis<sup>TM</sup>  
**Storage Conditions:** 4 – 25°C  
**Handling Conditions:** Keep sterile

**Date of Manufacture:** March 2023  
**Batch#:** 004  
**Lot #:** 0026  
**Expiration Date:** September 2024

### Testing Characteristic

### Test Result

### Lot

Compare Ct values between Reagent RVD + RT-qPCR to RNA extraction + RT-qPCR from stock sample

Observed Ct relative to RNA extraction:  
**+0.48**

Pass: +/- 5Ct:  
**PASS**

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

**PASS**

**Date of Quality Control Testing:** March 2023

This certificate of analysis verifies the above-mentioned lot of Reagent RVD with RVD Enhancer conforms to intended quality parameters on the indicated date above.

**For Research Use Only.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

  
Signature of Approving Manager