

Certificate of Analysis - Reagent RVD with RVD Enhancer

Product Name: Reagent RVD with RVD Enhancer
Catalog #: 78336001 / 78336025 / 783361000
Product Line: PCROpsisTM
Storage Conditions: 4 – 25°C
Handling Conditions: Keep sterile

Date of Manufacture: February 2022
Batch#: 005
Lot #: 0028
Expiration Date: August 2023

Testing Characteristic

Test Result

Lot

Compare Ct values between Reagent RVD + RT-qPCR to RNA extraction + RT-qPCR from stock sample

Observed Ct relative to RNA extraction:
-0.60

Pass: +/- 3 Ct:
PASS

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

PASS

Date of Quality Control Testing: April 2023

This certificate of analysis verifies the above-mentioned lot of Reagent RVD with RVD Enhancer conforms to intended quality parameters on the indicated date above.

For Research Use Only. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager


Signature of Approving Manager