



Entopsis, Inc.
 7600 NW 69th Ave
 Medley, FL 33166
 info@entopsis.com
 Ph#: 1-888-407-5070

Certificate of Analysis – Lysis Beads

<u>Product Name:</u>	Lysis Beads	<u>Date of Manufacture:</u>	May 2022
<u>Catalog #:</u>	597850 / 5970025	<u>Lot #:</u>	0002
<u>Product Line:</u>	PCRopsis™	<u>Expiration Date:</u>	May 2024
		<u>Storage Conditions:</u>	4 - 25°C
		<u>Handling Conditions:</u>	Keep sterile

Testing Characteristic	Test Result	Lot
Detection of <100 cfu / mL Candida albicans in urine	Lowest concentration detected (cfu / mL): 8.3	Pass: <100 PASS
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C	No Growth	PASS

Date of Quality Control Testing: May 2022

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

For In Vitro Diagnostic Use. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager