

Entopsis, Inc. 7600 NW 69th Ave Medley, FL 33166 info@entopsis.com Ph#: 1-888-407-5070

March 2023

Certificate of Analysis - Reagent RVD with RVD Enhancer

Date of Manufacture:

Catalog #:78336001Product Line:PCR $opsis^{TM}$ Storage Conditions: $4-25^{\circ}C$ Handling Conditions:Keep sterile	<u>Batch#</u> : <u>Lot #:</u> <u>Expiration Date:</u>	004 0033 September 2024
Testing Characteristic	Test Result	Lot
Compare Ct values between Reagent RVD + RT-qPCR to a Plasmid Positive Control	Observed Ct relative to RNA extraction: -1.15	Pass: +/- 3 Ct: PASS
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C	No Growth	PASS

Date of Quality Control Testing: August 2023

Product Name: Reagent RVD with RVD Enhancer

This certificate of analysis verifies the above-mentioned lot of Reagent RVD with RVD Enhancer conforms to intended quality parameters on the indicated date above.

For Research Use Only. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager