

## Certificate of Analysis - Reagent RVD with RVD Enhancer

**Product Name:** Reagent RVD with RVD Enhancer  
**Catalog #:** 78336001  
**Product Line:** PCROpsis<sup>TM</sup>  
**Storage Conditions:** 4 – 25°C  
**Handling Conditions:** Keep sterile

**Date of Manufacture:** February 2023  
**Batch#:** 002  
**Lot #:** 0038  
**Expiration Date:** August 2024

### Testing Characteristic

### Test Result

### Lot

Compare Ct values between Reagent RVD + RT-qPCR to a Plasmid Positive Control

Observed Ct relative to  
RNA extraction:  
**+0.99**

Pass: +/- 3 Ct:  
**PASS**

Test for microbial growth on blood agar  
plates during 72 hour incubation at 37°C

No Growth

**PASS**

**Date of Quality Control Testing:** January 2024

This certificate of analysis verifies the above-mentioned lot of Reagent RVD with RVD Enhancer conforms to intended quality parameters on the indicated date above.

**For Research Use Only.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

  
Signature of Approving Manager