



Entopsis, Inc.  
 7600 NW 69<sup>th</sup> Ave  
 Medley, FL 33166  
 info@entopsis.com  
 Ph#: 1-888-407-5070

## Certificate of Analysis - Reagent Buccal

**Product Name:** Reagent Buccal  
**Catalog #:** 282025 / 282100 / 282500  
**Product Line:** PCROpsis™

**Date of Manufacture:** April 2021  
**Lot #:** 0003  
**Expiration Date:** October 2022  
**Storage Conditions:** 4 - 25°C  
**Handling Conditions:** Keep sterile

### Testing Characteristic

### Test Result

### Lot

Compare Ct values between reagent + qPCR to DNA extraction + qPCR from stock sample

Observed Ct relative to DNA extraction:  
**+2.42**

Pass: +/- 5Ct:  
**PASS**

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

**PASS**

**Date of Quality Control Testing:** May 2021

This certificate of analysis verifies the above-mentioned lot of Reagent Buccal conforms to intended quality parameters on the indicated date above.

**For In Vitro Diagnostic Use.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

**Obdulio Piloto**

Name of Approving Manager



Signature of Approving Manager