



Entopsis, Inc.  
 7600 NW 69<sup>th</sup> Ave  
 Medley, FL 33166  
 info@entopsis.com  
 Ph#: 1-888-407-5070

## Certificate of Analysis - Reagent Buccal

**Product Name:** Reagent Buccal  
**Catalog #:** 282001 / 282025 / 282100  
**Product Line:** PCROpsis™

**Date of Manufacture:** February 2022  
**Lot #:** 0004  
**Expiration Date:** August 2023  
**Storage Conditions:** 4 - 25°C  
**Handling Conditions:** Keep sterile

Testing Characteristic	Test Result	Lot
Compare Ct values between reagent + qPCR to DNA extraction + qPCR from stock sample	Observed Ct relative to DNA extraction: <b>-2.6</b>	Pass: +/- 5Ct: <b>PASS</b>
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C	No Growth	<b>PASS</b>

**Date of Quality Control Testing:** February 2022

This certificate of analysis verifies the above-mentioned lot of Reagent Buccal conforms to intended quality parameters on the indicated date above.

**For In Vitro Diagnostic Use.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto  
 \_\_\_\_\_  
 Name of Approving Manager

  
 \_\_\_\_\_  
 Signature of Approving Manager