



## Certificate of Analysis – Lysis Beads

<b><u>Product Name:</u></b>	Lysis Beads	<b><u>Date of Manufacture:</u></b>	June 2022
<b><u>Catalog #:</u></b>	597850 / 597250	<b><u>Lot #:</u></b>	0005
	5970025 / 5970500	<b><u>Expiration Date:</u></b>	June 2024
<b><u>Product Line:</u></b>	PCRopsis™	<b><u>Storage Conditions:</u></b>	4 - 25°C
		<b><u>Handling Conditions:</u></b>	Keep sterile

### Testing Characteristic

### Test Result

### Lot

Detection of <100 cfu / mL Candida albicans in urine

Lowest concentration detected (cfu / mL):  
**65**

Pass: <100  
**PASS**

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

**PASS**

**Date of Quality Control Testing:** June 2022

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

**For In Vitro Diagnostic Use.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

\_\_\_\_\_  
Name of Approving Manager

  
\_\_\_\_\_  
Signature of Approving Manager