



## Certificate of Analysis – Activator

<b><u>Product Name:</u></b>	Activator	<b><u>Date of Manufacture:</u></b>	March 2023
<b><u>Catalog #:</u></b>	228001 / 228002	<b><u>Batch #:</u></b>	0001
<b><u>Product Line:</u></b>	PCRopsis™	<b><u>Lot #:</u></b>	0006
<b><u>Storage Conditions:</u></b>	10 – 25°C	<b><u>Expiration Date:</u></b>	September 2023
<b><u>Handling Conditions:</u></b>	Keep sterile		

### Testing Characteristic

### Test Result

### Lot

Compare Ct values between Reagent RVD-RT w/ Activator + RT-qPCR to RNA extraction + RT-qPCR from stock sample

Observed Ct relative to RNA extraction:  
**+0.48**

Pass: +/- 5Ct:  
**PASS**

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

**PASS**

**Date of Quality Control Testing:** March 2023

This certificate of analysis verifies the above-mentioned lot of Activator conforms to intended quality parameters on the indicated date above.

**For Research Use Only.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

\_\_\_\_\_  
Name of Approving Manager

\_\_\_\_\_  
Signature of Approving Manager