



Certificate of Analysis – Lysis Beads

Product Name: Lysis Beads
Catalog #: 597850 / 597250
5970025 / 5970500
Product Line: PCR_{opsis}TM
Storage Conditions: 20°C ~ 25°C
Handling Conditions: Keep sterile

Date of Manufacture: June 2022
Batch #: 001
Lot #: 0006
Expiration Date: June 2024

Testing Characteristic

Test Result

Lot

Detection of <100 cfu / mL Candida albicans in urine

Lowest concentration detected (cfu / mL):
50

Pass: <100
PASS

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

PASS

Date of Quality Control Testing: March 2023

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

For Research Use Only. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager