



Entopsis, Inc.
 7600 NW 69th Ave
 Medley, FL 33166
 info@entopsis.com
 Ph#: 1-888-407-5070

Certificate of Analysis - Reagent RVD-RT

| | | | |
|-----------------------------|---------------------------------|------------------------------------|---------------|
| <u>Product Name:</u> | Reagent RVD-RT | <u>Date of Manufacture:</u> | December 2021 |
| <u>Catalog #:</u> | 78378025 / 78378100 / 783781000 | <u>Lot #:</u> | 0006 |
| <u>Product Line:</u> | PCRopsis™ | <u>Expiration Date:</u> | June 2023 |
| | | <u>Storage Conditions:</u> | 4 - 25°C |
| | | <u>Handling Conditions:</u> | Keep sterile |

| Testing Characteristic | Test Result | Lot |
|---|---|-------------------------------|
| Compare Ct values between reagent + RT-qPCR to RNA extraction + RT-qPCR from stock sample | Observed Ct relative to RNA extraction: -4.53 | Pass: +/- 5Ct: PASS |
| Test for microbial growth on blood agar plates during 72 hour incubation at 37°C | No Growth | PASS |

Date of Quality Control Testing: January 2022

This certificate of analysis verifies the above-mentioned lot of Reagent RVD-RT conforms to intended quality parameters on the indicated date above.

For In Vitro Diagnostic Use. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager



Signature of Approving Manager