



Entopsis, Inc.
 7600 NW 69th Ave
 Medley, FL 33166
 info@entopsis.com
 Ph#: 1-888-407-5070

Certificate of Analysis – Activator

<u>Product Name:</u>	Activator	<u>Date of Manufacture:</u>	May 2023
<u>Catalog #:</u>	228001 / 228025	<u>Batch #:</u>	0002
<u>Product Line:</u>	PCRopsis™	<u>Lot #:</u>	0007
<u>Storage Conditions:</u>	10 – 25°C	<u>Expiration Date:</u>	November 2023
<u>Handling Conditions:</u>	Keep sterile		

Testing Characteristic

Test Result

Lot

Compare Ct values between Reagent RVD-RT w/ Activator + RT-qPCR to RNA extraction + RT-qPCR from stock sample

Observed Ct relative to RNA extraction:
+0.37

Pass: +/- 3Ct:
PASS

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

PASS

Date of Quality Control Testing: May 2023

This certificate of analysis verifies the above-mentioned lot of Activator conforms to intended quality parameters on the indicated date above.

For Research Use Only. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager