



Entopsis, Inc.  
 7600 NW 69<sup>th</sup> Ave  
 Medley, FL 33166  
 info@entopsis.com  
 Ph#: 1-888-407-5070

## Certificate of Analysis – Lysis Beads

<b><u>Product Name:</u></b>	Lysis Beads	<b><u>Date of Manufacture:</u></b>	July 2022
<b><u>Catalog #:</u></b>	5970025 / 5970500 / 5971000	<b><u>Lot #:</u></b>	0007
<b><u>Product Line:</u></b>	PCRopsis™	<b><u>Expiration Date:</u></b>	July 2024
		<b><u>Storage Conditions:</u></b>	4 - 25°C
		<b><u>Handling Conditions:</u></b>	Keep sterile

<b>Testing Characteristic</b>	<b>Test Result</b>	<b>Lot</b>
Detection of <100 cfu / mL Candida albicans in urine	Lowest concentration detected (cfu / mL): <b>65</b>	Pass: <100  <b>PASS</b>
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C	No Growth	<b>PASS</b>

**Date of Quality Control Testing:** July 2022

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

**For In Vitro Diagnostic Use.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

---

Name of Approving Manager

---

Signature of Approving Manager