



Entopsis, Inc.
 7600 NW 69th Ave
 Medley, FL 33166
 info@entopsis.com
 Ph#: 1-888-407-5070

Certificate of Analysis – Lysis Beads

<u>Product Name:</u>	Lysis Beads	<u>Date of Manufacture:</u>	April 2023
<u>Catalog #:</u>	5970010 / 5970025 / 5970500	<u>Lot #:</u>	0009
	/ 5971000	<u>Expiration Date:</u>	November 2024
<u>Product Line:</u>	PCRopsis™	<u>Storage Conditions:</u>	25°C
		<u>Handling Conditions:</u>	Keep sterile

Testing Characteristic

Test Result

Lot

Detection of <100 cfu / mL Candida albicans in urine

Lowest concentration detected (cfu / mL):
50

Pass: <100
PASS

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

PASS

Date of Quality Control Testing: May 2023

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

For In Vitro Diagnostic Use. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager