



## Certificate of Analysis – Lysis Beads

**Product Name:** Lysis Beads  
**Catalog #:** 5970010 / 5970025  
**Product Line:** PCRopsis™  
**Storage Conditions:** 25°C  
**Handling Conditions:** Keep sterile

**Date of Manufacture:** May 2023  
**Batch #:** 004  
**Lot #:** 0011  
**Expiration Date:** November 2024

### Testing Characteristic

### Test Result

### Lot

Detection of <100 cfu / mL Candida albicans in urine

Lowest concentration detected (cfu / mL):  
**50**

Pass: <100  
**PASS**

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

**PASS**

**Date of Quality Control Testing:** August 2023

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

**For In Vitro Diagnostic Use.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

\_\_\_\_\_  
Name of Approving Manager

  
\_\_\_\_\_  
Signature of Approving Manager