



Entopsis, Inc.
 7600 NW 69th Ave
 Medley, FL 33166
 info@entopsis.com
 Ph#: 1-888-407-5070

Certificate of Analysis - Reagent RVD-RT

<u>Product Name:</u>	Reagent RVD-RT	<u>Date of Manufacture:</u>	June 2022
<u>Catalog #:</u>	78378025 / 78378100 / 783781000	<u>Lot #:</u>	0011
<u>Product Line:</u>	PCRopsis™	<u>Expiration Date:</u>	June 2023
		<u>Storage Conditions:</u>	4 - 25°C
		<u>Handling Conditions:</u>	Keep sterile

Testing Characteristic	Test Result	Lot
Compare Ct values between reagent + RT-qPCR to RNA extraction + RT-qPCR from stock sample	Observed Ct relative to RNA extraction: -0.37	Pass: +/- 5Ct: PASS
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C	No Growth	PASS

Date of Quality Control Testing: June 2022

This certificate of analysis verifies the above-mentioned lot of Reagent RVD-RT conforms to intended quality parameters on the indicated date above.

For In Vitro Diagnostic Use. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager



Signature of Approving Manager