

Entopsis, Inc. 7600 NW 69<sup>th</sup> Ave Medley, FL 33166 info@entopsis.com Ph#: 1-888-407-5070

## **Certificate of Analysis – Lysis Beads**

Product Name: Catalog #: Product Line: Storage Conditions: Handling Conditions:	Lysis Beads 5970010 / 5970025 PCR <i>opsis</i> <sup>TM</sup> 25°C Keep sterile	<u>Date of Manufacture:</u> Batch #: <u>Lot #:</u> <u>Expiration Date:</u>	May 2023 004 0012 November 2024
<b>Testing Characteristic</b>		Test Result	Lot
Detection of <100 cfu / mL Candida albicans in urine		Lowest concentration detected (cfu / mL): 50	Pass: <100 PASS
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C		No Growth	PASS

Date of Quality Control Testing: August 2023

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

**For In Vitro Diagnostic Use.** If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager