



Entopsis, Inc.
 7600 NW 69th Ave
 Medley, FL 33166
 info@entopsis.com
 Ph#: 1-888-407-5070

Certificate of Analysis – Support

<u>Product Name:</u>	Support	<u>Date of Manufacture:</u>	September 2023
<u>Catalog #:</u>	787002	<u>Batch #:</u>	0003
<u>Product Line:</u>	PCRopsis™	<u>Lot #:</u>	0013
<u>Storage Conditions:</u>	4 – 25°C	<u>Expiration Date:</u>	September 2024
<u>Handling Conditions:</u>	Keep sterile		

Testing Characteristic	Test Result	Lot
Concentration of core ingredients	Relative amount: +2.97	Pass: +0 - 20 PASS
Test for microbial growth on blood agar plates during 72 hour incubation at 37°C	No Growth	PASS

Date of Quality Control Testing: October 2023

This certificate of analysis verifies the above-mentioned lot of Support conforms to intended quality parameters on the indicated date above.

For Research Use Only. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager