



Certificate of Analysis – Lysis Beads

Product Name: Lysis Beads
Catalog #: 5970010 / 5970025
Product Line: PCR^{opsis}TM
Storage Conditions: 25°C
Handling Conditions: Keep sterile

Date of Manufacture: October 2023
Batch #: 005
Lot #: 0015
Expiration Date: April 2025

Testing Characteristic

Test Result

Lot

Detection of <100 cfu / mL Candida albicans in urine

Lowest concentration detected (cfu / mL):
90

Pass: <100
PASS

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

PASS

Date of Quality Control Testing: December 2023

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

For In Vitro Diagnostic Use. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager



Signature of Approving Manager