

Certificate of Analysis – Support

<u>Product Name:</u>	Support	<u>Date of Manufacture:</u>	September 2023
<u>Catalog #:</u>	787003	<u>Batch #:</u>	0003
<u>Product Line:</u>	PCRopsis™	<u>Lot #:</u>	0015
<u>Storage Conditions:</u>	4 – 25°C	<u>Expiration Date:</u>	September 2024
<u>Handling Conditions:</u>	Keep sterile		

Testing Characteristic

Test Result

Lot

Concentration of core ingredients

Relative amount:

Pass: +0 - 20

+4.54

PASS

Test for microbial growth on blood agar plates during 72 hour incubation at 37°C

No Growth

PASS

Date of Quality Control Testing: February 2024

This certificate of analysis verifies the above-mentioned lot of Support conforms to intended quality parameters on the indicated date above.

For Research Use Only. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager


Signature of Approving Manager