

Entopsis, Inc. 7600 NW 69th Ave Medley, FL 33166 info@entopsis.com Ph#: 1-888-407-5070

Certificate of Analysis – Lysis Beads

<u>Product Name:</u> <u>Catalog #:</u>	Lysis Beads 5970010 / 5970025	<u>Date of Manufacture:</u> Batch #: Lot #:	October 2023 006 0018
Product Line: Storage Conditions: Handling Conditions:	PCR <i>opsis</i> ™ 25°C Keep sterile	Expiration Date:	April 2025
Testing Ch	aracteristic	Test Result	Lot
Detection of <100 cfu / mL Candida albicans in urine		Lowest concentration detected (cfu / mL):	Pass: <100 PASS
Test for microbial grov plates during 72 hour i	_	No Growth	PASS
Date of Quality Contro This certificate of analy parameters on the indica	vsis verifies the above-mention	3 oned lot of Lysis Beads conform	ns to intended quality
For In Vitro Diagnostic Use Services at 1(888) 407-5070		about this Certificate of Analysis, pleas	e contact Technical
Obdulio Piloto		Oll Re	
Name of Approv	ing Manager	Signature of Appro	ving Manager