

Entopsis, Inc. 7600 NW 69th Ave Medley, FL 33166 info@entopsis.com Ph#: 1-888-407-5070

Certificate of Analysis – Lysis Beads

Product Name: Catalog #: Product Line: Storage Conditions: Handling Conditions:	Lysis Beads 5970010 / 5970025 PCR <i>opsis</i> ™ 25°C Keep sterile	<u>Date of Manufacture:</u> Batch #: <u>Lot #:</u> <u>Expiration Date:</u>	October 2023 006 0019 April 2025
Testing Characteristic		Test Result	Lot
Detection of <100 cfu / mL Candida albicans in urine		Lowest concentration detected (cfu / mL): 60	Pass: <100 PASS
Test for microbial grov plates during 72 hour i		No Growth	PASS

Date of Quality Control Testing: December 2023

This certificate of analysis verifies the above-mentioned lot of Lysis Beads conforms to intended quality parameters on the indicated date above.

For In Vitro Diagnostic Use. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1(888) 407-5070 or info@entopsis.com.

Obdulio Piloto

Name of Approving Manager

Signature of Approving Manager